



Attestation of Conformity

No. E8A 054966 0007 Rev. 01

Holder of Certificate: Gastron Co.,Ltd.

23, Gunpocheomdansaneop 1-ro Gunpo-si, Gyeonggi-do 15881 REPUBLIC OF KOREA

Name of Object:

Electrical equ. for measurement, control and

laboratory use (GAS DETECTOR)

Model(s):

GHD-100A

Description of

Rated input voltage:

230-240 V a.c.

Object:

Rated frequency:

50 Hz

Rated input power:

1.5 W

Tested according to: EN 50270:2015 EN 61000-6-2:2005

EN 61000-6-3:2007/A1:2011

EN 61000-3-2:2014 EN 61000-3-3:2013

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.:

CTK-2019-00133

Date,

2019-02-08

(Byung-Soo Kang)

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU Declaration of conformity the required CE marking can be affixed on the product. That Declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed. TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany