



IECEx Certificate of Conformity

INTERNATIONAL ELECTROTECHNICAL COMMISSION IEC Certification Scheme for Explosive Atmospheres

for rules and details of the IECEx Scheme visit www.iecex.com

Certificate No.: **IECEx SIR 08.0003** issue No.: **6**

Status: **Current**

Date of Issue: **2014-03-31** Page 1 of 4

Applicant: **Gastron Co. Limited**
18-8, Dogeumdanji 1-Gil
Palgogi-Dong
Sangrok-Gu
Ansan-Si, Gyeonggi-Do
Korea, Republic of

Certificate history:
Issue No. 6 (2014-3-31)
Issue No. 5 (2014-3-20)
Issue No. 4 (2014-1-22)
Issue No. 3 (2013-4-17)
Issue No. 2 (2010-7-15)
Issue No. 1 (2009-2-5)
Issue No. 0 (2008-11-18)

Electrical Apparatus: **Type GTD-2000Ex, GTD-2000Tx & GIR-3000 Gas Detectors**
Optional accessory:

Type of Protection: **Flameproof**


Marking:	GTD-2000Ex	GTD-2000Tx and GIR-3000		
	Ex d IIC Gb T (see below)	Ex d IIC Gb T (see below)		
	T Class	Ambient temperature	T Class	Ambient temperature
	T6	Ta = -40°C to +60°C	T6	Ta = -40°C to +60°C
	T5	Ta = -40°C to +60°C	T4	Ta = -40°C to +80°C
	T4	Ta = -40°C to +80°C	T4	Ta = -40°C to +80°C

Approved for issue on behalf of the IECEx Certification Body: **C Ellaby**

Position: **Deputy Certification Manager**

Signature:
(for printed version)

Date:


2014-03-31

1. This certificate and schedule may only be reproduced in full.
2. This certificate is not transferable and remains the property of the issuing body.
3. The Status and authenticity of this certificate may be verified by visiting the [Official IECEx Website](http://www.iecex.com).

Certificate issued by:

SIRA Certification Service
Rake Lane
Eccleston
Chester
CH4 9JN
United Kingdom

sira
CERTIFICATION



IECEx Certificate of Conformity

Certificate No.: IECEx SIR 08.0003

Date of Issue: 2014-03-31

Issue No.: 6

Page 2 of 4

Manufacturer: **Gastron Co. Limited**
18-8,Dogeumdanji 1-Gil
Palgogi-Dong
Sangrok-Gu
Ansan-Si,Gyeonggi-Do
Korea, Republic of

Additional Manufacturing location
(s):

This certificate is issued as verification that a sample(s), representative of production, was assessed and tested and found to comply with the IEC Standard list below and that the manufacturer's quality system, relating to the Ex products covered by this certificate, was assessed and found to comply with the IECEx Quality system requirements. This certificate is granted subject to the conditions as set out in IECEx Scheme Rules, IECEx 02 and Operational Documents as amended.

STANDARDS:

The electrical apparatus and any acceptable variations to it specified in the schedule of this certificate and the identified documents, was found to comply with the following standards:

IEC 60079-0 : 2007-10 Edition: 5	Explosive atmospheres - Part 0:Equipment - General requirements
IEC 60079-1 : 2007-04 Edition: 6	Explosive atmospheres - Part 1: Equipment protection by flameproof enclosures "d"
IEC 60079-29-1 : 2007 Edition: 1	Explosive Atmospheres - Part 29-1: Gas Detectors - Performance requirements of detectors for flammable gases

*This Certificate **does not** indicate compliance with electrical safety and performance requirements other than those expressly included in the Standards listed above.*

TEST & ASSESSMENT REPORTS:

A sample(s) of the equipment listed has successfully met the examination and test requirements as recorded in

Test Report:

GB/SIR/ExTR08.0141/00
GB/SIR/ExTR13.0065/00
GB/SIR/ExTR14.0079/00

GB/SIR/ExTR09.0011/00
GB/SIR/ExTR14.0010/00

GB/SIR/ExTR10.0171/00
GB/SIR/ExTR14.0069/00

Quality Assessment Report:

GB/SIR/QAR08.0021/00
GB/SIR/QAR08.0021/03

GB/SIR/QAR08.0021/01
GB/SIR/QAR08.0021/04

GB/SIR/QAR08.0021/02



IECEx Certificate of Conformity

Certificate No.: IECEx SIR 08.0003

Date of Issue: 2014-03-31

Issue No.: 6

Page 3 of 4

Schedule

EQUIPMENT:

Equipment and systems covered by this certificate are as follows:

The GTD-2000Ex Combustible Gas Detector consists of an enclosure, Type GDH-1000 (IECEX SIR 08.0112U), fitted with a Type GSA860Ex gas sensor (IECEX SIR 08.0111U). The enclosure is fitted with associated circuitry, termination facilities and a display visible through the GDH-1000 window. Cable entry is by means of the suitably certified cable entry device fitted into the GDH-1000 threaded aperture. The device is rated up to 24 V and 200 mA.

The GTD-2000Tx Toxic Gas Detector consists of an enclosure, Type GDH-1000 (IECEX 08.0112U), fitted with a Type GSA860Tx gas sensor (IECEX SIR 08.0111U). The enclosure is fitted with associated circuitry, termination facilities and a display visible through the GDH-1000 window. Cable entry is by means of the suitably certified cable entry device fitted into the GDH-1000 threaded aperture. The device is rated up to 24 V and 150 mA.

CONDITIONS OF CERTIFICATION: NO

Empty box for conditions of certification.



IECEx Certificate of Conformity

Certificate No.: IECEx SIR 08.0003

Date of Issue: 2014-03-31

Issue No.: 6

Page 4 of 4

DETAILS OF CERTIFICATE CHANGES (for issues 1 and above):

Issue 1 – this Issue introduced the following change:									
1.	The addition of the GIR 3000 Gas Detector consisting an enclosure, Type GDH-2000 (IECEx 08.0112U), fitted with a Type GSA920 Infra Red Combustible Gas Sensor (IECEx SIR 08.0111U). The enclosure is fitted with associated circuitry, termination facilities and a display visible through the GDH-2000 window. Cable entry is by means of the suitably certified cable entry device fitted into the GDH-2000 threaded aperture. The device is rated up to 400 mA @ 24 Vdc.								
Issue 2 – this Issue introduced the following change:									
1.	It was recorded that performance testing has been successfully carried out on the GTD-2000Ex and GIR-3000 Gas Detectors, as a consequence the list of standards was updated to include EN 61779-1:1998 Ed. 1 and EN 61779-4:1998 Ed. 1.								
Issue 3 – this Issue introduced the following changes:									
1.	It was recognised that these Gas Detectors are now available as a stainless steel option; this is due to the fact that the component approved enclosures used in their construction can now be made from this material.								
2.	The change associated with Issue 2 was editorially amended to clarify those Gas Detectors that have been subjected to performance testing.								
Issue 4 – this Issue introduced the following changes:									
1.	Following appropriate assessment the Type GTD 2000Ex Gas Detector was approved for use in the following ambient temperature ranges with temperature classes as specified below:								
	<table border="1"> <thead> <tr> <th>Temperature Class</th> <th>Ambient temperature</th> </tr> </thead> <tbody> <tr> <td>T6</td> <td>Ta = -40°C to +60°C</td> </tr> <tr> <td>T5</td> <td>Ta = -40°C to +60°C</td> </tr> <tr> <td>T4</td> <td>Ta = -40°C to +80°C</td> </tr> </tbody> </table>	Temperature Class	Ambient temperature	T6	Ta = -40°C to +60°C	T5	Ta = -40°C to +60°C	T4	Ta = -40°C to +80°C
Temperature Class	Ambient temperature								
T6	Ta = -40°C to +60°C								
T5	Ta = -40°C to +60°C								
T4	Ta = -40°C to +80°C								
2.	The applicants address was changed from 75-10 Palgok 2-Dong, Sangrok-Gu, Ansan City, Kyunggi-Do to 18-8, Dogeumdanji 1-Gil, Palgogi-Dong, Sangrok-Gu, Ansan-Si, Gyeonggi-Do.								
Issue 5 – this Issue introduced the following change:									
1.	Following appropriate assessment, the Type GTD-2000Tx and GIR-3000 Gas Detectors were approved for use in the following ambient temperature ranges with temperature classes as specified below: The marking was amended accordingly.								
	<table border="1"> <thead> <tr> <th>Temperature Class</th> <th>Associated ambient temperature (°C)</th> </tr> </thead> <tbody> <tr> <td>T6</td> <td>Ta = -40°C to +60°C</td> </tr> <tr> <td>T4</td> <td>Ta = -40°C to +80°C</td> </tr> </tbody> </table>	Temperature Class	Associated ambient temperature (°C)	T6	Ta = -40°C to +60°C	T4	Ta = -40°C to +80°C		
Temperature Class	Associated ambient temperature (°C)								
T6	Ta = -40°C to +60°C								
T4	Ta = -40°C to +80°C								
Issue 6 – this Issue introduced the following change:									
1.	Following appropriate assessment to demonstrate compliance with the requirements of the IEC 60079 series of standards, the documents previously listed in Issue 5 of the certificate, IEC 61779-1:1998 and IEC 61779-4:1998 were replaced by IEC 60079-29-1:2007.								